Hands-on learning and implementing using LabVIEW™ for undergraduates in 13 weeks

Alex See, PhD

**Monash University Malaysia, School of Engineering and Science,
No. 2 Jalan Kolej, Bandar Sunway, 46150, PJ, Selangor Darul Ehsan, Malaysia
email: alex.see@engsci.monash.edu.my**

Abstract

Second year Mechatronics undergraduates, in the year 2003 at Monash University Malaysia (MUM) were taking a subject module known as Project and Practice. Students were exposed to National Instrument's LabVIEW™ software and hardware for the first time. They were required to complete a given mechatronically oriented project in only 13 weeks. A team of 3 students was formed to design, develop and test a prototype solar-tracker, which involves two plane-parallel solar panels with position tracking system. This solar tracker built by these students consisted of *two-axes*. This system was capable of tracking the direction as well as the degree of inclination of the position of the sun throughout the day. The main objective of this project design was to always align the photovoltaic flat type solar panels towards the direction of maximum light intensity from the sun. The purpose was to obtain maximum solar irradiance and hence maximizing solar power extraction. As this was a laboratory prototype system, preliminary testing was conducted indoors. A 100-Watt light bulb from a conventional table lamp was used to generate sufficient light illumination. The sensing of light intensity was achieved by using four PIN photodiodes operating in photovoltaic mode. Prior to this project, the students had no previous experience in LabVIEW™ software at all.

This paper reports that National Instruments LabVIEW™ and MAX (Measurement And automation eXplorer) software may be employed as useful tools for practical teaching and learning for undergraduates in Mechatronics. The useful functions in LabVIEW™ allowed this group of students to appreciate the fundamentals of motion control for stepper motors, Analog I/O (input/output) and fundamentals of data acquisition. In this Mechatronics oriented subject, students learned about the various skills and acquired knowledge through a hands-on approach.

The students were provided with a motion controller card (model PCI-7334) and Universal Motion Interface (model UMI-7764) from National Instruments for this project work.

1. Introduction

Solar panels find many practical applications such as thermal energy storage systems, electric power generating systems, aerospace industries¹ etc. The conversion of solar radiation into electrical energy by means of solar cells has been developed as part of satellite and space travel technology². The average solar energy intercepted by conventional stationary solar panels, during the course of the day is not fully optimized. Obviously, this is due to the static solar panel's position/placement, which hinders it from exposure to the sun throughout the course of the day¹. Extensive work had been performed for the determination of the total solar irradiation on a flat surface at any orientation and tilt^{1, 3, 4, 5, 6}. The study conducted by Koay and Bari⁵ had revealed that there could be an increased in 37% of the available solar energy for useful work, if there was ability to track the sun continuously during daylight hours during the day. To date, at the time of this writing, a personal computer based tracking system for solar collector had been recently reported by Chong et al⁷. Chong et al system utilized an automatic interval tracking system, which do not use any photo sensors but deriving estimation function for tracking purpose. In the work done by Chong et al, the estimation of the solar radiation required the user to enter tracking interval, the day and the latitude in question. The software would formulate the required orientation of the solar collector, and position the solar collector to maximum solar irradiance using stepper motors. This method of solar tracking is a more economical method and eliminates the use of sensors and feedback system, though its main drawback is that it cannot offer real-time tracking.

This paper describes the work produced by a team of 3 students. They were put together for a hands-on learning experience in Mechatronics engineering. The rationale for hands-on learning in engineering had been widely studied by many researchers and in one particular study conducted by Ferguson and Hegarty⁸, they carried out an experiment and investigated on how college students learned the mechanics of pulley systems. Their study showed that students with hands-on equipment demonstrated a significant enhancement in their abilities to solve practical problems related to the real-engineering world, as compared to students who learned with diagrams. This group of 3 students designed and developed a prototype real-time solar tracking system, which comprises of two plane-parallel solar panels. Although sun tracker systems are commercially available in the market, they can be quite costly depending on their level of sophistication in tracking. A graphical programming language, LabVIEWTM (Laboratory Virtual Instrument Engineering Workbench) had been chosen for this implementation. Students were exposed to LabVIEWTM software in particular the FlexMotion functions in LabVIEW version 6.1. The students used a motion controller card (model PCI-7334) from National Instruments to control a two-axes stepper motor system in only 13 weeks.

2.0 GSE2800-project and practise module

Every second year mechatronics students at Monash University Malaysia are required to enroll for a module known as the Project and Practice. This particular mechatronics subject is targeted at students with main intentions of allowing students to learn engineering design and skill building through a hands-on approach. This module is compulsory for the students

in their second year, second semester. Students will develop skills and knowledge by investigating a given mechatronically oriented project. This module comprises of 1-hour lecture and 3-hours of laboratory work per week. This module of the course lasts for 13 weeks in the second semester of every year. Each year, the actual available duration for students to work on the project is only *11 weeks*. The first week of the second semester is normally catered for students' orientation and administration matters. The final 13th week is reserved for project presentation and report writing. Therefore, the total number of scheduled laboratory time is only 33 hours in every second semester of each year.

There were a total of 12 students taking this Project and Practice module in the last semester 2003. They were divided into four groups, and each group was presented with a project that was different from the others. The intention of this paper is to present the prototype real-time solar tracker project designed by a group of three students. The students' gain of knowledge and appreciation for hands-on learning will be illustrated.

3.0 System setup and description

In this section, description of the system setup would be highlighted. The students spent about 4 weeks to design the mechanical structure of this prototype solar tracker. They used the AutoCAD software for the mechanical structure design. 2 weeks were used for designing and constructing the photodiodes sensing circuits. The rest of the time was mainly spent on programming, debugging and testing.

3.1 System block diagram and structural design

The system block diagram for this prototype design is shown in figure 3.1.1.

Figure 3.1.1 System block diagram of the prototype solar tracker system incorporated using National Instrument motion controller card (model PCI-7334), Universal Motion Interface (UMI-7764), LabVIEWTM and auxillary components are depicted above.

Figure 3.1.2 Structure overview of the prototype solar tracker system is shown above. Basic components are shown. It includes two stepper motors and drivers for horizontal and vertical movement. The photodiodes, which are used as light sensors are positioned in a special pyramid shape mounted onto a flat perspex sheet. Two solar panels are mounted as shown in the above figure.

Figure 3.1.3 depicts the completed 2-axes solar tracker setup at the laboratory. Front view showing photodiodes and two solar panels. The pyramid structured photodiodes sensing circuit and the two flat plane solar panels are shown mounted on a flat piece of supporting material made of perspex.

The model of the two stepper motors and drivers in used were VEXTA CSK series 2-phase type. The operating power supplies to these stepper motors were either 24 or 36 VDC. The stepper motor outputs were configured as the Clockwise mode. This stepper motors had angular resolution of 1.8° and would require 200 steps to complete a revolution.

The photodiodes were mounted in a pyramid formation as shown in figure 3.1.4. They were arranged in such a fashion so that the opposite pairs of the photodiodes were used for tracking both the horizontal and vertical position of the light source. The light sensing circuit used is basically a (I-to-V) current to voltage converter using a low power LM158 operational amplifier. There were two planes (i.e. horizontal and vertical) and 4 photodiodes were used. The students built 4 similar (I-to-V) converter circuits. The current produced by the photodiode is proportional to the light intensity received from external source (i.e. light bulb), the voltage from the sensing circuit would increase or decrease according. The 4 voltage signals from the outputs of the 4 sensing circuits were connected to the 4-quadrature inputs of the UMI-7764.

The photodiodes were isolated from one another as a result of mounting them on the pyramid support structure. The intention was to blind one sensor from the other. This had led to such a design and prototyping of a pyramid shaped base that would maximize sensitivity. A large voltage differentiate would be detected if the light source was not placed directly perpendicular to the adjacent pairs of left and right photodiodes as well as

the top and bottom photodiodes. The differentiated voltages were used as feedback signals. These signals were used by the LabVIEW program to command the movement of the two stepper motors.

Figure 3.1.4 shows a sketch of the front view of the photodiodes in this pyramid formation.

4.0 LabVIEW as the choice of software package to utilize

National Instruments LabVIEWTM software is a graphical programming language that utilizes icon instead of using lines text codes, for e.g. C++. Unlike others, it is a programming language that is user friendly and easy to learn. LabVIEWTM was developed by National Instrument in 1986⁹. There were a few reasons why LabVIEWTM was chosen as the choice of software package to use for the undergraduate learning. One of the main reasons was that LabVIEWTM is easy to use and the learning period is substantially short for someone who has already had some experience in basic programming. These students had some experience with Java programming prior to taking this project and practise module. The other motivating factor for the choice of LabVIEWTM was that there are extensive documentation available, to name a few e.g.¹⁰⁻¹³ in LabVIEWTM. There is also a huge amount of ready available VI (Virtual Instrumentations) that students can utilize and exploit. In a survey paper by Nesimi Ertugrul¹⁴, he highlighted that LabVIEWTM software is a good tool for teaching/learning laboratory based work and that the virtual instrumentation approach is open to further improvements and development. This may increase the student participation and enthusiasm in the process.

4.1 MAX (Measurement & Automation eXplorer)

Measurement & Automation explorer (MAX) is the National Instruments (NI) configuration utility that is used to configure the motion controller in used for this project. Appropriate settings and initialization had to be performed before turning the stepper motors in this project.

Figure 4.1.1 shows an example of the MAX configuration utilities displaying the PCI-7334 motion controller hardware that is available in the PC. Various parameters had to be set appropriately before the actual moving of the stepper motors in this work.

4.2 PCI 7334 and UMI 7764

The low cost NI PCI-7334 motion controller card is suitable for this application because it supports up to 4-axes stepper motor control applications. For this project, the students used only two axes and the motion of the stepper motors were programmed using LabVIEW™. The Universal Motion Interface was used for the interconnections to third party stepper motor power drivers¹⁰. It provided a comprehensive wiring and connection point for motion control and feedback signals. A single cable from the motion controller to the UMI carried all the input and output signals for all the 4 axes. For this work, the four quadrature analog inputs (i.e. 0 to 5 VDC) of the UMI-7764 were used to accept individually, the outputs of the four I-to-V sensing circuits.

5.0 LabView programming details

This section illustrates how the students were introduced to LabVIEW™. The students were each given three copies of the following manuals namely, NI-Motion Control user manual¹⁰, LabVIEW™ Basics 1 Introduction Course Manual¹² and LabVIEW™ User Manual¹³. They were also given the LabVIEW™ evaluation CD-ROM so that they were able to try this software at their own free time.

5.1 Students exposure to labview for the first time

These students had no knowledge or experience in LabVIEW™ prior to this course. During the course, students would have a chance to understand the nature of LabVIEW™ programming through a scheduled 4 hours intensive lecture. During the lecture, LabVIEW™ environment, including windows, menus, tools, front panels, block diagrams, dataflow programming and various tools palette available in LabVIEW™ software were shown to students to enhance their learning.

Generally, students were taught about the commonly used structures in programming like the 'WHILE', 'FOR', 'CASE' structures, 'SEQUENCE' that are available in the function palette at the block diagram level. Adequate time was required by the instructor to explain to the students the main functionality and features in the LabVIEW™ environment. As examples, the following 'SEQUENCE' and 'WHILE' Loop structures are briefly explained.

Case Structure. It has one or more subdiagrams, or cases, exactly one of which executes when the structure executes. Only one case is visible at a time and the only one case can be executed. There is a case selector label at the top of the case structure. It consists of the name of the selector value that corresponding to the case in the center with a decrement and increment arrows beside. The structure has a TRUE case and a FALSE case. If the selector terminal is an integer, string or enumerated type value, the structure can have arbitrarily many cases. If the data type of the wire connected to the selector terminal of a case structure has been changed, the structure will automatically convert the case selector values to the new data type when possible.

While Loop Structure. The while loop is similar in the text-based programming to execute a subdiagram continuously until a condition is met. There is a conditional terminal or an input terminal on the right end corner of the loop. It only receives a specific Boolean value. The default behavior of the conditional terminal is Continue If True. The loop will be continuously executed as long as the input value is TRUE. In other words, the loop will be terminated until the conditional terminal receives FALSE. The behavior can be changed by right-click on the terminal or the border of the While loop and selecting Stop If True. When it has been changed to this behavior, the

loop executes its subdiagram until the terminal receives a TRUE value. Because the VI checks the conditional terminal at the end of each iteration, the while loop always executes at least one time. The VI is broken if the terminal is not connected. There is an output terminal called iteration terminal on the left of the loop shown. It shows the number of completed iteration. The count always starts at zero. During the first iteration, the iteration number returns zero.

In fact, it was impossible for the instructor to explain and cover all the functionality in LabVIEW™ to the students in such a short span of allocated time. Students were encouraged to spend time on their own to explore in more details about the functionality of LabVIEW™ and its features.

5.2 Programming details

The student's design of their software flow chart is depicted as shown in figure 5.2.1. A brief description of the program flow is as follows:

- This LabVIEW™ program was supposed to read in 4 analog voltage values from the output of the 4 individual photodiode-operational amplifier sensing circuits via the 4 analog input channels of UMI-7764.
- Comparisons of voltage readings were performed in the software for the two pairs of sensors. The vertical opposite pair of sensors was meant for vertical plane tracking. The horizontal opposite pair of sensors was meant for horizontal plane tracking.
- If values differ from one another in each pair, the LabVIEW™ program would generate commands to turn the stepper motor one step (i.e. 1.8°) at a time until voltage readings were the same in each pair. Once equilibrium was achieved, the solar panels would be aligned towards the direction of maximum light intensity from the light source.

Figure 5.2.1. Flow chart of the software program design

The figure 5.2.2 shows the LabVIEW™ FlexMotion VIs that is readily available for used in motion control applications.

Figure 5.2.2 (a) to (c) show the function palette, Motion & Vision sub-palette and various available FlexMotion VIs that can be used for the purpose of motion control applications.

Initially, the students had developed a simple LabVIEWTM program for turning the stepper motor and later resetting the target position of the stepper motor as shown in figure 5.2.3. Students had to explore the various parameters such as target position, velocity, acceleration etc.

Figure 5.2.3 A simple program that is able to move the stepper motor.

The students were able to write simple program for acquiring analog voltage values from the I-to-V photodiode/op-amp sensing circuits. The LabVIEWTM diagram is shown in figure 5.2.4.

Figure 5.2.4 shows the A/D read Vi available for used for reading in analog values from 0 to 5 VDC.

The following diagram shown in figure 5.2.5 is only a partial listing of the complete block diagram. This program was tested and it had successfully demonstrated that this solar tracker was able to track the maximum intensity of a light source. The resolution of tracking would actually be the resolution of the stepper motors, which in this case is 1.8° for both the horizontal and vertical position.

Figure 5.2.5 A partial listing of the fully developed student's LabVIEW™ program, which only shows two ADC input channels 1 and 2. The other two input channels 3 & 4 are not shown here for simplicity.

The students had used a separate PCI-6036E DAQ card and wrote a simple LabVIEW program to acquire the open circuit voltage of the series connected solar panels. The measured open circuit voltage profile is shown in the figure 5.2.6. A single channel data acquisition block diagram for open circuit voltage measurement is shown in figure 5.2.7.

Initially, up to about 8 sec, the solar tracker was at rest, as the light was turned off at that time. The table lamp was positioned at about 50 cm away from the system and was turned on at about 9 sec. The solar tracker started to respond due to the sudden strong light intensity and this can be seen as a result of the increasing open circuit voltage from the solar panels. The tracker continued to track both its vertical and the horizontal position of the stationary light source and maintained a steady condition after about 25 s in time.

Figure 5.2.6 Open circuit voltage profile of the series connected solar panels versus tracking time. The sampling interval is 0.5 sec. The tracking system appears to maintain steady condition after about 25s.

Figure 5.2.7 A single channel data acquisition block diagram for open circuit voltage measurement

6.0 Student response to practical hands-on learning

Initially, students were quite intimidated and they found LabVIEW™ software quite daunting and overwhelming due to unfamiliarity. However, after getting accustomed to LabVIEW™, the three students were generally impressed by its functionality, in particular the Motion control – FlexMotion VIs. Students were able to control the stepper motors rotation precisely. The overall responses obtained from the students were positive and they were generally satisfied with the outcome of their project work. The students commented that the available documentation and examples in LabVIEW™ were clear and easy to follow. They gave a remark *“Through this project, we have learned various skills and gained knowledge in particular to engineering designs, development and implementation. We also learn about working as a team and effective communication with lecturer and our peers”*.

7.0 Conclusion remarks

In this paper, the hands-on learning and implementing using LabVIEW™ for undergraduates had been demonstrated successfully. Students were able to design, develop and test a prototype 2-axes solar tracker system using LabVIEW™ software in only a short span of time (i.e. 13 weeks). The students had *no previous* experiences in LabVIEW™ before and it has to be noted that rapid prototyping and software development by the students were made possible through the use of LabVIEW™. This project had helped the students gain much knowledge in different disciplines of engineering. Students had engaged themselves actively in the area of mechanical, electrical and electronics circuit designs and also in software designs. Students had applied their basic skills and knowledge in this mechatronically oriented project work. The skill building and knowledge acquisition by the students through this hands-on approach had been achieved with much success.

The use of LabVIEW™ in this project and practise subject may be an effective way for hands-on learning for this group of mechatronics students.

8.0 Acknowledgement

The author gratefully acknowledges Monash University Malaysia, in particular to Associate Professor Maki Habib for his support. Special thanks goes to Mr Paneer and Mr Shahrul for their laboratory technical support.

Bibliography

1. B. Koyuncu and K. Balasubramanian, "A microprocessor controlled automatic sun tracker", *IEEE trans. on Consumer Electronics*, Vol. 37, No. 4, pp 913 – 917, (1991)
2. D. Yogi Goswami, Frank Kreith and Jan F. Kreider, *Principles of Solar Engineering*, Second edition, Taylor and Francis publishing, ISBN 1-56032-714-6
3. M. Tiris and C. Tiris, "Optimum collector slope and model evaluation: Case study for Gebze, turkey", *J. Energy and conversion Mgmt*, Vol. 39, No. 3/4, pp 167-172, (1998)
4. S. Bari, "Optimum slope angle and orientation of solar collectors for different periods of possible utilization", *J. Energy and conversion Mgmt*, Vol. 41, No 8, pp 855-860, (2000)
5. C.W. Koay and S. Bari, "Intermittent tracking of flat plate collectors", *Proceeding of the World Renewable energy congress, Malaysia*, pp 79-82, (1999)
6. Abdul-Jabbar N. Khalifa and Salman S. AL-Mutawalli, "Effect of two-axes sun tracking on the performance of compound parabolic concentrators", *J. Energy and conversion Mgmt*, Vol. 39, No 10, pp 1073-1079, (1998)
7. L.C Chong, Z. Samad, S. Bari, W.C. Au and A. Ameen, "Development of a personal computer based tracking system for solar collector", *Proceeding of the Intl. Symposium on Renewable energy*, Kuala Lumpur, Malaysia, late paper, (2003)
8. E.L. Ferguson and M. Hegarty, "Learning with real machines or diagrams: application of knowledge to real-world problems", *Cognition and instruction*, Vol. 13, No. 1, pp 129-160, (1995)
9. Beyon, J. Y., *LabVIEW programming, data acquisition and analysis*, Prentice Hall PTR (Upper Saddle River, NJ) 2001.
10. Motion Control NI-Motion™ User manual – National Instruments, May 2003 Edition, Part Number 323367A-01
11. Hands-On Introduction to LabVIEW™ Graphical Development Environment Seminar – National Instruments, May 2003 Edition, Part Number 350150J-01

12. LabVIEW™ Basics 1 Introduction course manual – National Instruments, Course software version 7.0 June 2003 Edition, Part Number 320628L-01

13. LabVIEW™ User manual – National Instruments, November 2001 Edition, Part Number 320999D-01

14. Nesimi Ertugrul, “Towards Virtual Laboratories: a Survey of LabVIEW-based teaching/learning tools and future trends”, *Intl. J. of Engineering Education*, Vol 16, No. 3, pp 171-181, (2000).

Biography

Alex See obtained his Bachelor degree (Hons) in Electrical and Electronics Engineering from University of Leicester, UK in 1998. He was given a fee waiver scholarship at University of Leicester to pursue his PhD in the area of high voltage engineering immediately after his graduation. He had obtained his PhD in the year 2001 and thereafter he worked with the Defense Science Organisation (Singapore) as an engineer for one year. He is currently a lecturer at Monash University at Malaysia campus.